

M183

I.V. Trainer with Indicator



SAKAMOTO MODEL
CORPORATION

User's Manual (131025)

— I N D E X —

• Introduction/Table of Contents	1
• Purpose/Content	2
• Cautions on Use	3
• Instructions for Use	4
• Storage Instructions	5

! Note

Confirm the main unit and accessories directly after the product arrives. If there is any defect or breakage, inform the dealer or Sakamoto Model.

株式会社坂本モデル



Introduction

Thank you for purchasing our product, “I.V. Trainer with Indicator.”

Read this user’s manual carefully in order to use the product for a long time appropriately.

This product is closely checked before factory shipment. However, when this product arrives, open it soon, and check if anything on the main unit and accessories are missing or broken.

Table of contents

Introduction／Table of Contents.....	1
Purpose／Content.....	2
Cautions on Use.....	3
Instructions for Use.....	4
Storage Instructions.....	5

Purpose

The following training (evaluation) can be performed.

- 1) Intravenous injection training
- 2) Blood sampling training

Content

- 1) Simulated skin set - Qty. 1
- 2) Pump - Qty. 1
- 3) Attachment belt - Qty. 2
- 4) Exchangeable blood vessel - Qty. 1
- 5) AA size alkaline batteries - Qty. 4



Cautions on Use

- 1) Be careful that the connector in the middle of the tube does not disconnect accidentally during use.
- 2) Use new needle tips when performing injection training with this product. Otherwise, it may cause the skin to deteriorate.
- 3) If there are frequent blood vessel leaks and/or the deteriorating condition is severe, replace the blood vessel with a new one.
- 4) Never inject the needle in an upright direction nor use excessive force that may pierce all the way through. Doing so is extremely dangerous.
- 5) When using this product for training purposes, conduct the training in pairs with one person playing the role of the patient and wearing this product.
- 6) Be careful not to get the top of the pump and mechanical parts wet.
- 7) If a defect or problem with the product is discovered during use, stop using the product immediately and contact Sakamoto Model.
- 8) This product is a simulator capable of supporting advanced medical training. However, Sakamoto Model does not guarantee the same results during actual treatment.
- 9) Please note that the specifications of this product may be changed without prior notice in order to make improvements to the product.

Instructions for Use

- 1) Take out the training kit, and place it on a table with a non-slick surface for training.
- 2) Get some water ready.

- 3) Open the top of the pump and fill it up to the line with water.
(Always perform these steps before Step 4.)



- 4) Connect the cable and the tubes that come out of the skin pieces to the pump.



- 5) Open the pump cover, and insert the 4 AA size alkaline batteries.
- 6) Turn on the switches located on the sides of the pump.




- 7) Begin training.
- 8) Attach the belt to the Velcro strip on the back of the skin pieces, and fit it onto the arm of the person playing the patient in order to practice training.
- 9) Refer to the storage instructions for putting this product away.

Blood Vessel Exchange Procedure



First, perform steps 1 through 5 in the “Storage Instructions” before replacement, and then perform the following.

- 1) Open the skin set, and disconnect the 4 locations noted in the diagram. Then, attach following the same procedure when attaching a new blood vessel.



Storage Instructions

- 1) Turn the switch off.
- 2) Close the white clasp.
- 3) Disconnect the LAN cable and the 2 tubes that are attached to the pump.
- 4) Discharge the water that is inside the pump.
- 5) Open the white clasp to discharge the water that is inside the tube.
- 6) Even though the product is not exposed to direct sunlight when stored, the remaining water should evaporate naturally.
- 7) Remove the batteries and store each part.
- 8) Avoid storing the product in direct sunlight or in locations with high temperature and humidity.



SAKAMOTO MODEL
C O R P O R A T I O N

Head office

34,Higashitakagicho,Shimogamo
Sakyo-ku, KYOTO 606-0865, JAPAN

Phone: +81-75-701-1135

Fax: +81-75-722-5638

E-mail: info@sakamoto-model.com

URL: <http://sakamoto-model.co.jp>

Overseas Sales Department

1-9-29,Senbahigashi,Minoh City,Osaka
562-0035,JAPAN

Phone: +81-72-729-8350

Fax: +81-72-729-8385

E-mail: trade@sakamoto-model.com